

вместе с колебаниями находящихя в ней МНЧ. При наложении на биологическую среду постоянного магнитного поля магнитные моменты МНЧ ориентируются в направлении этого поля и синфазно колеблются с частотой ультразвука. Колебания МНЧ вызывают появление в окружающем пространстве вторичного переменного магнитного поля. Это поле регистрируется чувствительным магнитометром, находящимся вблизи участка биологической среды.

Выводы. Из изложенного следует, что магнитоакустические и акустомагнитные методы обнаружения МНЧ являются предпочтительными.

Список использованных источников

1. Shigang W., Shunqi Z., Ren M, Tao Y., Zhipeng L. A study of acoustic source generation mechanism of Magneto acoustic Tomography // Computerized Medical Imaging and Graphics.- 2014.- Vol.38.- P.42-48.
2. Xu Y, He B. Magneto acoustic tomography with magnetic induction // Physics in Medicine and Biology.- 2005.- Vol.50.- P.5175-5187.
3. Mariappan L., Shao Q., Jiang C., Kai Y., Ashkenazi S. et al. Magneto acoustic tomography with short pulsed magnetic field for in-vivo imaging of magnetic iron oxide nanoparticles // Nanomedicine: Nanotechnology, Biology, and Medicine.- 2016.- Vol.12.- P.689-699.
4. Avrunin O.G., Bondarenko I.S. Kharkiv National University of Radio Electronics. ICBEM and RGC conference in Aachen (в печати).

OPTOELECTRONIC METHODS FOR ANALYSIS OF MICROCIRCULATORY DISORDERS IN INFLAMMATORY PROCESSES IN THE MAXILLOFACIAL REGION

¹Pavlov S.V., ¹Barylo A.S., ¹Kozlovska T.I., ²Kravchuk P.O.

¹*Vinnitsa National Technical University;*

²*Vinnitsa National Medical University*

Introduction. In the pathogenesis of odontogenic inflammatory diseases, an important role plays disruption of blood circulation in the tissues of the maxillofacial region as consider many authors (V.M.Uvarov, M.M.Solovyev, T.M.Alehova). These statements are based on the topographic anatomical, morphological studies and on data of studying the blood coagulation. However, all of these methods allow only indirectly define the state of the regional blood flow. Therefore, use of such actual

methods as rheoplethismography and photoplethismography is very perspective. These methods allow to study the pathogenesis of periodontal disease, periodontitis, mumps and other pathological processes in the maxillofacial region.

Materials and methods. In recent years, the photoplethismographic methods for registration of biosignals are introduced in the practice of functional diagnostics. These methods based on irradiation the area of the tissue of the biological object by infrared beam and recording, transmitted through the tissue or reflected from it radiation by optoelectronic sensor. The use of optoelectronic sensors has provided new opportunities for the diagnosis of the state of the cardiovascular system.

The main advantage of this method is the possibility of measuring the microcirculation level almost anywhere in the body surface, that allows to use the different modifications of optical devices to solving a number of specific tasks related to the study of indicators of local blood flow (diagnosis of microcirculation in spinal motion segments, assessing the state of the microcirculation of the lower legs in the disease of the systemic lupus erythematosus, vascular state assessment in the maxillofacial region in inflammatory processes, analysis of hemodynamic blood flow in pulse diagnosis, and etc.). Furthermore, depending on the particular method of measurement may be estimated such parameters as the biomedical total concentration of hemoglobin, total blood oxygen saturation (the degree of oxygen saturation), total bilirubin concentration. Photoplethismographic curve shows the phase changes of blood filling of the peripheral vascular in inflammatory processes, respectively cardiac cycle.

To improve the reliability of received photoplethismographic information uses *a priori* information, including the physical characteristics of the measurement object, the mathematical relation between the measured values, the data about the spectral composition of informative components and noise, and also basic biophysical characteristics of the controlled object. The engineered optoelectronic complex allows to diagnose the state of blood vessels by express method at different stages of the disease and to fix the degree of microcirculation and hemodynamic disturbances in some areas by comparing the received signals.

This device is a converter for displaying and comparing the transformed biomedical signals. It is possible to connect the device to PC via a coupling unit. This significantly expands such functionality of the device as recording of biomedical signals to the archive, their preprocessing (scaling, filtering),

comparison and correlation analysis and building the charts and graphs on the PC.

Results. As a result of researches that carried out at the Department of Maxillofacial Surgery of the Vinnitsa National Medical University, it was examined 30 patients (21 men and 9 women aged 20 to 60 years) with odontogenic abscesses and phlegmon. Treatment of patients included removal of the causative tooth, opening cellulitis or abscess, ozone therapy with silver ions.

Photoplethysmographic studies were carried out on the optoelectronic diagnostic complex for analysis of microcirculatory disorders. The optical radiation was directed to the biological tissue at a distance of 10 mm from the edge of the surgical wound or intended cut. As a control point was symmetric point of study. Photoplethysmographic signals (PPGS) were registered in the inflammatory focus and in the control point before the operation, and on the third and fifth day after operation. There were recorded 180 PPGS and the obtained data were processed by designed program «Wosteo».

Discussing of the results. During the analysis of obtained data it was set that photoplethysmographic signal of phlegmon and abscess significantly differed of control.

Before surgery, the level of blood filling (photoplethysmographic index) in the inflammatory focus is significantly increased in 2.5-3.5 times. State of venous outflow was sharply deteriorated, that was appeared in the change of form of decaying limb (it that has become more prominent – 73.3%). Dicrotic jag was less expressed and shifted to the top of the catacrotism (70%). In 22 cases is noted the appearance of additional venous waveform. The blood flow velocity is reduced due to the further deterioration of the venous outflow. In 76.6% of cases there were additional dicrotic waves, 30% were marked small additional jags on catacrotism.

At the 5-6 days after surgery blood flow was improved in all indicators. Additional waves were disappeared in 46.6%, and were weakened in 53.4%. However, in the case of PPGS of inflammatory focus the difference from control point was remained.

Conclusion. Using of the photoplethysmographic method is allows to assess accurately the level of blood-filling in inflammatory processes, thus, this method has the such positive properties: noninvasive, high degree of sensitivity and reliability, ease of study. Application of this method allows dentists to determine accurately the effectiveness of the treatment; to

specify the duration of the rehabilitation period; to identify various vascular disorders in fractures of the jaw; to evaluate the effectiveness of local anesthesia (anesthesia causes vasospasm, and reducing of the amplitude of the signal is the feature of the effectiveness of anesthesia); to apply this method to plastic surgery and transplantation.

References

1. Pavlov S.V. Optoelectronic devices for diagnosis of peripheral circulation with high reliability / S.V.Pavlov, T.I.Kozlovska, V.B.Vasilenko.- Vinnitsa: NTB, 2014.- 140 p.
2. Pavlov S.V. Physical principles of biomedical optics / S.V.Pavlov, V.P.Kozhemiako, P.F.Kolesnik, T.I.Kozlovska, V.P.Dumenko.- Vinnitsa: NTB, 2010.- 155 p.
3. Sander S.V. Laser photoplethysmography in integrated evaluation of collateral circulation of lower extremities / S.V.Sander, T.I.Kozlovska, V.B.Vasilenko, V.S.Pavlov, A.Y.Klapouschak, P.Kisala, R.S.Romaniuk // Proc. SPIE 9816, Optical Fibers and Their Applications.- 2015, 98161K; doi: 10.1117/12.2229042.

РОЗРОБКА ОПТИЧНОГО КАСКАДУ СВІТЛОДІОДНОЇ РУКОЯТКИ ДЛЯ ПРИЛАДУ «LIKA-LED»

Холін В.В., Івасенко В.І., Чепурна О.М., Павлов С.В., Тітова Н.В., Рева А.В., Петраш М.Т, Биченко В.М., Петрушко Ю.А.

Прилад «LIKA-LED» призначений для проведення наукових досліджень. Він складається з електронного блоку і набору рукояток. В якості випромінюючого елемента рукоятки використано світлодіод. Конструктивне виконання приладу дозволяє підключати до одного електронного блоку одночасно три незалежні одна від одної рукоятки для роботи з декількома довжинами хвиль.

Метою даної роботи є розробка оптичної системи світлодіодної рукоятки приладу «LIKA-LED».

Світлодіодна рукоятка складається з джерела випромінювання (світлодіода), оптичного каскаду та електричної частини.

Одною з задач є підбір оптичних елементів, які повинні забезпечити: мінімальний кут розходження променів; однорідний розподіл щільності потужності в зображенні; максимально паралельний світлодіодний потік випромінювання. Тому, для розробки оптичного каскаду рукоятки було підібрано і використано різного типу лінзи. Асферична для зменшення кута розкриття променів та сферична для концентрації випромінювання в більш вузький потік світла.