



Materials
50-th International Scientific
and Practical Conference

APPLICATION OF LASERS IN MEDICINE AND BIOLOGY

22–25 May 2019
Kharkiv, Ukraine

MINISTRY OF EDUCATION AND SCIENCE OF UKRAINE

V. N. KARAZIN KHARKIV NATIONAL UNIVERSITY

**Materials
50th Anniversary International
Scientific and Practical Conference**

Application of Lasers in Medicine and Biology

22–25 May 2019

Kharkiv

Kharkiv
2019

INTELLECTUALIZED SYSTEM OF 2D- MULLER –MATRIX TOMOGRAPHY FOR DIAGNOSTICS OF FRACTIONAL LAYERS OF BIOLOGICAL TISSUE

Zabolotna N.I., Pavlov S.V., Ronald Rovira

Vinnitsia National Technical University, Ukraine

Topicality of the work. The further progress of the methods of polarize-sensible optical coherent tomography may be associated with the development of new methods for analysis and processing the polarize nonhomogeneous images of the real multi-layer biological tissues (BT), which form the physiological organs. The issues of using the laser polarimetry for diagnosing the oncologic changes in the under-surface layers, shielded by other layers of biological tissues of this or the other human organ remain poorly researched.

Our work is aimed at the experimental research of the possibilities of polarized reconstruction (tomography) of the coordinate distributions of Muller matrix elements of the under-surface layers of the smooth muscle, which is shielded by the layer of the conjunctive tissue, cervix of the uterus in the system of 2D-Muller –matrix tomography.

Methods and research materials. The architecture of the system contains the measuring (radiator, collimator, block of polarizing scanning, objecting block, projecting block, 6 polarized filters, block for registering and two-measuring data discretization), which forms the Muller matrix of the BT structure, as well as PC with special software. The latter realizes the algorithmic reproduction of the “orientational” and “phasic” 2D-distributions (tomograms) of the birefringence crystals in the net of polarizationally reproduced under-surface layer and their analysis (statistical, correlation, fractal).

Optically-thin (weakening factor $\tau \leq 0.1$) histological cryomicroscopic sections of the healthy and pathologically changed (cancerous state - dysplasia) tissue of the neck of uterus were used as the research material

As the diagnostic parameters of the Muller matrix of the neck of uterus tissue there had been chosen two of its diagonal elements γ_{22} and γ_{44} , which characterize correspondingly the “orientational” and “phasic” properties of the net of biological crystals of the layer of smooth muscle.

Results and their discussion. The table presents the values of statistical, correlation and spectral moments, which characterize the coordinate, autocorrelation and fractal distributions of the directly measured

(optically-thin layers of the healthy tissue of the smooth muscle and in the state of dysplasia) and polarized reproduced (two-layers structure“smooth muscle– connecting tissue) of matrix elements γ_{22}^* ($m \times n$) and γ_{44}^* ($m \times n$).

Table

Y_{ik}	Y_{22}				Y_{44}			
	norm		pathology		norm		pathology	
$M^{(1)}$	0,31	0,27*	0,38	0,33*	0,25	0,21*	0,19	0,25*
$M^{(2)}$	0,18	0,2*	0,21	0,24*	0,14	0,1*	0,08	0,11*
$M^{(3)}$	0,31	0,36*	0,49	0,57*	0,29	0,19*	0,77	0,56*
$M^{(4)}$	3,35	3,91*	4,19,	4,71*	1,64	1,14*	6,19	4,99*
$Q^{(2)}$	0,22	0,24*	0,18	0,22*	0,18	0,23*	0,14	0,17*
$Q^{(4)}$	1,14	1,42*	1,43	1,59*	0,97	0,74*	1,85	1,11*
$J^{(1)}$	0,62	0,57*	0,55	0,63*	0,59	0,68*	0,66	0,72*
$J^{(2)}$	0,14	0,19*	0,16	0,22*	0,17	0,12*	0,34	0,28*
$J^{(3)}$	0,19	0,21*	0,27	0,24*	0,24	0,15*	0,41	0,29*
$J^{(4)}$	0,42	0,32*	0,48	0,36*	0,32	0,22*	0,57	0,36*

The analysis of the data shows that method of Muller matrix reproduction of the polarized properties of the optic – anisotropic structure of the smooth muscle in the multi-layer tissue of the uterus is of high efficiency – the maximal differences between the experimental parameters and polarizely reproduced do not exceed 35% - 45%.

It had been shown that the most diagnostically efficient for the differentiation of the optic properties of the healthy and oncologically changed layer of the tissue of the smooth muscle of the uterus are: 2-d, 3-d and 4-th statistic moments of coordinate distributions γ_{44}^* ($m \times n$) – differences between them are within 1,8 (M_2) - 4 (M_4) times; the 4-th correlation moment- differences between them are up to 2 times; 22-d, 3-d and 4-th spectral moments - differences between them are up to 1,65 – 2 times.

Conclusion. For Muller matrix images of the polarized reproduced orientation elements γ_{22}^* ($m \times n$) of the layer of the tissue of the smooth muscle of the uterus, the diagnostic efficiency of the method of polarized modulation tomography appeared to be insignificant, - the difference between the statistical $M_{i=1;2;3;4}$, corellational $Q_{i=2;4}$ and spectral $J_{i=1;2;3;4}$ parameters of the both groups of samples do not exceed 15% - 35%.

References

1. Timchenko, Leonid I.; Pavlov, Sergii V.; Kokryatskaya, Natalia I.; et al. Bio-inspired approach to multistage image processing: Photonics Applications In Astronomy, Communications, Industry, And High Energy Physics Experiments 2017 Book Series: Proceedings of SPIE Volume: 10445 Article Number: UNSP 104453M Published: 2017.
2. Wojcik, W; Cakala, S; Kotyra, A; et al. Analysis of the operation of an electrooptical Pockels effect: Technology And Applications Of Light Guides Book Series: Proceedings Of The Society Of Photo-Optical Instrumentation Engineers (SPIE) Volume: 3189 Pages: 110-121 Published: 1997.
3. Maciejewski, Marcin; Surtel, Wojciech; Wojcik, Waldemar; et al. Telemedical systems for home monitoring of patients with chronic conditions in rural environment: Annals Of Agricultural And Environmental Medicine Volume: 21 Issue: 1 Pages: 167-173 Published: 2014.
4. Smolarz, Andrzej; Gromaszek, Konrad; Wojcik, Waldemar; et al. : Optical methods and artificial intelligence in diagnostics of industrial pulverized coal burner: Przegląd Elektrotechniczny Volume: 88 Issue: 9B Pages: 259-261 Published: 2012.
5. Kotyra, Andrzej; Wojcik, Waldemar; Gromaszek, Konrad; et al. Assessment of biomass-coal co-combustion on the basis of flame image. Przegląd Elektrotechniczny Volume: 88 Issue: 11B Pages: 295-297 Published: 2012.
6. Methods of Processing Video Polarimetry Information Based on Least-Squares and Fourier Analysis // RH Rovira, SV Pavlov, OS Kaminski, MM Bayas - Middle-East Journal of Scientific Research, T. 16 (9), 1201-1204 2013. – P.1201-1204.
7. Zabolotna N. I.; Pavlov S. V.; Ushenko A. G.; Karachevtsev A. O.; Savich V. O., et al. System of the phase tomography of optically anisotropic polycrystalline films of biological fluids, *Proc. SPIE* 9166, Biosensing and Nanomedicine VII, 916616 (August 27, 2014)
8. Zabolotna N. I.; Pavlov S. V.; Ushenko A. G.; Sobko O. V., and Savich V. O. Multivariate system of polarization tomography of biological crystals birefringence networks, *Proc. SPIE* 9166, Biosensing and Nanomedicine VII, 916615 (August 27, 2014); doi:10.1117/12.2061105.
9. Roman N. Kvyetnyy, Olga Yu. Sofina, Alla V. Lozun, and etc. “Modification of fractal coding algorithm by a combination of modern technologies and parallel computations”, *Proc. SPIE* 9816, Optical Fibers and Their Applications 2015, 98161R (17 December 2015).
10. Roman Kvyetnyy, Yuriy Bunyak, Olga Sofina, and etc. “Blur recognition using second fundamental form of image surface”, *Proc. SPIE* 9816, Optical Fibers and Their Applications 2015, 98161A (17 December 2015).
11. Roman N. Kvyetnyy, Olexander N. Romanyuk, Evgenii O. Titarchuk, and etc. “Usage of the hybrid encryption in a cloud instant messages exchange system“, *Proc. SPIE* 10031, Photonics Applications in Astronomy, Communications, Industry, and High-Energy Physics Experiments 2016, 100314S (28 September 2016).